

Food and Drug Administration 10903 New Hampshire Avenue Document Control Center – WO66-G609 Silver Spring, MD 20993-0002

September 3, 2014

Whip Mix Corporation Mr. John P. Waters Regulatory Compliance Officer 361 Farmington Avenue Louisville KY, 40217

Re: K140877

Trade/Device Name: Vericore Zirconia Blanks

Regulation Number: 21 CFR 872.6660

Regulation Name: Porcelain powder for clinical use

Regulatory Class: II Product Code: EIH Dated: April 1, 2014

Received: August 6, 2014

Dear Mr. Waters:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the <u>Federal Register</u>.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm.

Sincerely yours,

Mary S. Runner -S

Erin I. Keith, M.S.
Director
Division of Anesthesiology, General Hospital,
Respiratory, Infection Control and Dental Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Indications For Use		
510(k) Number (if known):K14087	7	
Indications for Use: Vericore zirconia blanks are made from CAD/CAM or manual milling machines, fabricate;	n pre-sintered zirconium	
 Zirconia Substructures Restorations (Including inlays, Crown Framework in the Anter Bridge Framework in the Anter 	rior and Posterior region	
Prescription Use X (21 CFR 801 Subpart D)	AND/OR	Over-The-Counter Use(21 CRR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Page __of__

K140877

510(K) SUMMARY

This summary of 510(k) safety and effectiveness information is being submitted in accordance with the requirements of 21 CFR 807.92

Date Prepared: 4/1/2014

1. APPLICANT

Whip Mix Corporation 361 Farmington Avenue Louisville, KY 40217

PHONE:

502-634-5357

FAX:

502-634-4512

EMAIL:

jwaters@whipmix.com

2. SUBMITTER and CONTACT

John P. Waters Regulatory Compliance Officer & Official Correspondent for Whip Mix Corporation 361 Farmington Avenue Louisville, KY 40217

PHONE:

502-634-5357

FAX:

502-634-4512

EMAIL:

jwaters@whipmix.com

DATE:

4/1/2014

3. **DEVICE NAME**

Vericore Zirconia Blanks

4. COMMON OR USUAL NAME AND CLASSIFICATION

Powder, Porcelain

Regulation Number: 872.6660

Product Code: EIH Classification: Class II

5. PREDICATE DEVICE INFORMATION

White Peaks Copran ZR/Origin ZR (K092496)

6. **DEVICE DESCRIPTION**

Vericore Zirconia Blanks is a device made from pre-pressed and pre-sintered zirconia powder. It is available in various shades, shapes, and sizes to accommodate the customers CAD/CAM or manual milling equipment/strategies and is not machine specific. Vericore Zirconia Blanks are made from a biocompatible zirconia powder. It is intended to be used by professionals for the fabrication of dental restorations.

7. INTENDED USE

Vericore zirconia blanks are made from pre-sintered zirconium dioxide intended to be used with many CAD/CAM or manual milling machines. Vericore zirconia blanks are biocompatible and designed to fabricate;

- Zirconia Substructures
- Restorations (Including inlays, onlays, and veneers)
- Crown Framework in the Anterior and Posterior regions
- Bridge Framework in the Anterior and Posterior regions

8. SUBSTANTIAL EQUIVALENCE WITH PREDICATE DEVICES

Whip Mix Corporation	White Peaks Dental Systems
Vericore Zirconia Blanks	Copran Zr/Origin YZ
Class II Device	Class II device
510(k) Pending	510(k)092496
Product Code EIH	Product Code EIH
Regulation Number- 872.6660	Regulation Number- 872.6660
Material- Biocompatible zirconia powder	Material- Biocompatible zirconia powder
manufactured by Tosoh Corporation	manufactured by Tosoh Corporation
Indications For Use	Indications For Use
Vericore zirconia blanks are made from pre-	Copran Zr! Origin YZ Zirconia blanks are
sintered zirconium dioxide intended to be used	presintered blanks for CAD CAM or manual
with many CAD/CAM or manual milling machines.	milling, made from biocompatible, tetragonal
Vericore zirconia blanks are biocompatible and	and polycrystalline zirconiumdioxyde. Milling
designed to fabricate;	blanks designed for;
-	- Crown frameworks in the anterior and
- Zirconia Substructures	posterior areas
- Restorations (Including inlays, onlays, and	- Bridge frameworks in the anterior and
veneers)	posterior areas
- Crown Framework in the Anterior and	- Primary conical crowns and telescopic
Posterior regions	crowns
 Bridge Framework in the Anterior and 	- Canti-levered bridges with a max. of
Posterior regions	one pontic having a premolar width

- Inlays, Onlays, Veneers

9. BENCH TESTING

Flexural strength, thermal expansion, and chemical solubility tests were performed in accordance with ISO 6872 and all tests passed. Non-clinical testing for density was performed as well and the results are recorded in the proposed labeling.

10. BIOCOMPATIBILITY

The product is biocompatible because the predicate device was tested in accordance with ISO 10993-10, 10993-3, and 10993-5. No further biocompatibility tests are necessary because Whip Mix Vericore Blanks are made from the same powder as the predicate.

11. SAFETY AND EFFECTIVENESS CONCLUSION

Vericore® Zirconia Blanks are substantially equivalent to White Peaks ZR material in safety and effectiveness when used in accordance with the instructions for use. Both have similar Indications for Use, use the same raw material, and are biocompatible.